

BIDFORM FOR THE PUBLIC BIDDING OF:

UPLB-RES-025-9-17 SUPPLY AND DELIVERY OF BRANDED AND BRAND NEW LABORATORY AND SCIENTIFIC EQUIPMENT FOR BIOTECH

DATE OF PREBID CONFERENCE:

Friday, September 8, 2017 2:30 PM

DEADLINE OF SUBMISSION OF BIDS:

Friday, September 22, 2017 1:15 PM

OPENING OF BIDS:

Will follow shortly after the deadline of the submission of bids

LOT #	ITEM #	REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS (write down detailed offered specifications)	REFERENCES (include supporting documents)	COMMENTS/ REMARKS (clarify inclusion/ exclusion)	EVALUATION
1	1	BRANDED AND BRANDNEW BENCHTOP SHAKING INCUBATOR, 3.3 cu.ft (92L) 230 volts/60Hz incubator shaker				
		- agency approval UL or CSA or EU approved				
		- oscillation rate 30-400 rpm (1rpm increments)				
		- temperature uniformity +/-05°C at 37°C				
		- temperature range Ambient +8°C to 60°C				
		- platform dimension 17 x 18 x .75" (44 x 45 x 2cm)				
		- capacity 3.3 cu.ft (92L) with over temperature alarm				
		- exterior dimension WxDxH in. (cm) 22 x 25.5 x 28 (55.9 x 64.8x71.2)				
		- interior dimension WxDxH in. (cm) 19 x 18 x 16.5 (48.2 x 45.7x41.9)				
		- capacity 3.3 cu.ft (92L) with over temperature alarm				
		-nominal Voltage 50/60Hz (V) 220-240				
		- maximum energy consumption (W) 840				
		- maximum Load 18lb (8kg)				
		Includes accessories:				
		a) 10 pc - flask holder, stainless steel 250ml (max cap: 25)				
		b) 5 pc - flask holder, stainless steel, 500ml (max cap.: 5)				
		c) 1 pc - test tube holder rack				
		d) 220 AVR KVA				
		Features:				
		Microprocessor Control; Load Flexibility; Platform included Independent controls for temperature, timer and shaker				
		Warranty: 1 year on parts and services				
		Delivery period: 30 Calendar days upon receipt of P.O.				

BIDFORM FOR THE PUBLIC BIDDING OF:

UPLB-RES-025-9-17 SUPPLY AND DELIVERY OF BRANDED AND BRAND NEW LABORATORY AND SCIENTIFIC EQUIPMENT FOR BIOTECH

DATE OF PREBID CONFERENCE:

Friday, September 8, 2017 2:30 PM

DEADLINE OF SUBMISSION OF BIDS:

Friday, September 22, 2017 1:15 PM

OPENING OF BIDS:

Will follow shortly after the deadline of the submission of bids

LOT #	ITEM #	REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS (write down detailed offered specifications)	REFERENCES (include supporting documents)	COMMENTS/ REMARKS (clarify inclusion/ exclusion)	EVALUATION
1	2	BRANDED & BRAND NEW HORIZONTAL ELECTROPHORESIS SYSTEM				
		<i>Specifications:</i>				
		- output voltage: 35V, 50V or 100V				
		- timer: 099 min or cont.				
		- safety: Lid interlock				
		- max. Buffer (ml): 230				
		- gel capacity: One: 10.5 x 6cm				
		- Two: 5 x 6cm				
		- weight: 1 lbs/0.45 kg				
		- dimensions: 7.5 x 5.1 x 2.1 in, 119 x 13 x 5.5 cm				
		- electrical: Autovolt 115V or 230V, 50/60Hz				
		Includes:				
		- gel box, power supply, two 10.5 x 6cm and four 5x6cm gel trays, casting stand for each size gel tray and two double sided combs for each gel tray. Also inclusive of 22/12 double sided comb (multichannel capable)				
		- warranty: 1 year on parts and services				
		- delivery period: 30 Calendar days upon receipt of P.O.				

BIDFORM FOR THE PUBLIC BIDDING OF:

UPLB-RES-025-9-17 SUPPLY AND DELIVERY OF BRANDED AND BRAND NEW LABORATORY AND SCIENTIFIC EQUIPMENT FOR BIOTECH

DATE OF PREBID CONFERENCE:

Friday, September 8, 2017 2:30 PM

DEADLINE OF SUBMISSION OF BIDS:

Friday, September 22, 2017 1:15 PM

OPENING OF BIDS:

Will follow shortly after the deadline of the submission of bids

LOT #	ITEM #	REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS (write down detailed offered specifications)	REFERENCES (include supporting documents)	COMMENTS/ REMARKS (clarify inclusion/ exclusion)	EVALUATION
1	3	BRANDED & BRAND NEW REFRIGERATED CENTRIFUGE				
		- speed: 500-15,000 rpm in 1rpm steps				
		- max Rcf: 22,000 G				
		- timer: 0.999 minutes + hold in 1 second steps				
		- dimensions: HWD: 35x40x62cm				
		- weight: 45 kilos without rotor				
		- temp: -9 to 40C PID controlled accuracy and control to 0.5C				
		Inclusive of:				
		a) Rotor for 1.5-2 ml tubes, 24 placers, tube size: 11x50mm tube size, Minimum Speed: 500rpm, Max. speed: 15,000rpm, rcf:22,000xg, autoclavable at 121°C with high domed seal polycarbonate lid				
		b) Rotor for 50ml tubes, 8 placers, 30x120mm tube size, Minimum speed 500rpm, Max speed: 6000rpm, RCF: 4800xg, autoclavable at 121°C				
		c) AVR 5KVA				
		Warranty: 1 year on parts and services				
		Delivery period: 30 Calendar days upon receipt of P.O.				
PR NO. 16110						

BIDFORM FOR THE PUBLIC BIDDING OF:

UPLB-RES-025-9-17 SUPPLY AND DELIVERY OF BRANDED AND BRAND NEW LABORATORY AND SCIENTIFIC EQUIPMENT FOR BIOTECH

DATE OF PREBID CONFERENCE:

Friday, September 8, 2017 2:30 PM

DEADLINE OF SUBMISSION OF BIDS:

Friday, September 22, 2017 1:15 PM

OPENING OF BIDS:

Will follow shortly after the deadline of the submission of bids

LOT #	ITEM #	REQUIRED SPECIFICATIONS	OFFERED SPECIFICATIONS (write down detailed offered specifications)	REFERENCES (include supporting documents)	COMMENTS/ REMARKS (clarify inclusion/ exclusion)	EVALUATION
-------	--------	-------------------------	--	--	---	------------

Note: Please fill-up applicable columns.

Signature of Representative: _____

Name of Representative: _____

End-user:

Position: _____

DR. EDWIN P. ALCANTARA

Signature over Printed Name

Company: _____

Address: _____

Telephone/ Fax: _____

Email: _____