

BIDFORM FOR THE PUBLIC BIDDING OF:

UPLB-RES-004-3-17 SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF REACTION VESSEL SET-UP WITH COMPLETE ACCESSORIES - REBID

DATE OF PRE-BID CONFERENCE:

Friday, May 19, 2017 1:15 PM

DEADLINE OF SUBMISSION OF BIDS:

Friday, June 2, 2017 1:00 PM

OPENING OF BIDS:

Will follow shortly after the deadline of the submission of bids

LOT #	ITEM #	REQUIRED SECIFICATIONS	OFFERED SPECIFICATIONS (write down detailed offered specifications)	REFERENCES (include supporting documents)	COMMENTS/ REMARKS (clarify inclusion/ exclusion)	EVALUATION
		*Supply, Delivery, Installation and Commissioning of Reaction Vessel Set-up with Complete Accessories to be delivered at Institute of Chemistry, UP College, Los Baños, Laguna				
1	1	*BRANDED AND BRAND NEW REACTION VESSEL SET-UP WITH COMPLETE ACCESSORIES				
		Must contain the following:				
		20-L capacity borosilicate glass jacketed reaction vessel r/b PTFE stopcock;				
		O-ring seal, coated silicone, DN200;				
		5-necked reaction vessel lid with 29/32 central socket;				
		metal adapter, DN25 to M16, straight;				
		PTFE seal for DN25;				
		hose connector 1.5 meters;				
		stand support with wheels and auxiliary desk;				
		O-ring seal, coated silicone, for valve, 20mm;				
		support device DN200;				
		mechanical stir bearing rod 10mm M-29/32;				
		double cardan drive, SS rod 10mm;				
		SS shaft, anchor type, DN200, 10mm;				
		double nut for overhead stirrer;				
		7L-capacity refrigerated circulator (-40 to 200 deg.C), 240V, 50Hz;				
		pH meter probe, external Pt100 probe with cable length of approx. 3 meters, with tool kit for nuts and bolts				
		The entire set-up must be checked for system proper functioning while in the manufacturing facility (prior to delivery).				
		The unit must be suitable for use with highly corrosive substances such as inorganic acids and bases.				
		The unit must be compatible with an existing IKA overhead stirrer.				

BIDFORM FOR THE PUBLIC BIDDING OF:

UPLB-RES-004-3-17 SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF REACTION VESSEL SET-UP WITH COMPLETE ACCESSORIES - REBID

DATE OF PRE-BID CONFERENCE:

Friday, May 19, 2017 1:15 PM

DEADLINE OF SUBMISSION OF BIDS:

Friday, June 2, 2017 1:00 PM

OPENING OF BIDS:

Will follow shortly after the deadline of the submission of bids

LOT #	ITEM #	REQUIRED SECIFICATIONS	OFFERED SPECIFICATIONS (write down detailed offered specifications)	REFERENCES (include supporting documents)	COMMENTS/ REMARKS (clarify inclusion/ exclusion)	EVALUATION
		Warranty terms: One (1) year for parts and services upon delivery and installation of the unit				
		Delivery: 30CD UPON RECEIPT OF NTP				
		Payment Term: 30 Days upon contract completion and acceptance				

Note: Please fill-up applicable columns.

Signature of Representative: _____

Name of Representative: _____

End-user:

Position: _____

JOHN KRIS F. BARTOLOME

Signature over Printed Name

Company: _____

Address: _____

Telephone/ Fax: _____

Email: _____