



UNIVERSITY OF THE PHILIPPINES

LOS BANOS

Los Banos, IV-A

VAT Reg. TIN: 000-864-006-00004

Request for Quotation/ Bid Form (Technical Specifications)
INSTITUTE OF CHEMISTRY

UPLB BAC SECRETARIAT
BY: J DATE: 6-3-22
JUN 08 2022 - 5PM
UPLB-RQ- 6-212-22-RES
DEADLINE OF SUBMISSION

UPLB-RQ-

DEADLINE OF SUBMISSION:

Suppliers Name:

Date
Fund Code: 824832C
MOP: SMALL VALUE PROCUREMENT
Contact No: 0995 679 6217
Contact Person: Ma. Jonah Girl H. Amoranto

Please quote your lowest price on the item/s listed below, subject to the General Conditions below.

Note:

- 1. Bidders shall provide correct and accurate information required in this form. All entries must be typewritten or in print and properly accomplished. Do not leave blank entries, put N/A for not applicable.
2. Price quotation/s to be denominated in Philippine Peso shall include all taxes, duties, and/ or levies payable.
3. Bidders must indicate the BRAND and MODEL NUMBER for equipment and its accessories or peripherals. Evidence shall be in the form of manufacturer's un-amended sale literature, unconditional statement of specification and compliance issued by the manufacturer and sample.
4. Quotation through fax/email is acceptable. Winning bidder shall submit original signed RQ before issuance of Purchase order (P.O.).
5. Quotations exceeding the Appoved Budget for Contract shall be rejected.
6. Documentary requirements per Memorandum No. 03 Series of 2017 shall be attached upon submission of the quotation
7. Others:

Table with 10 columns: ITEM No., GENERAL NAME OF THE ITEM, REQUIRED SPECIFICATIONS, UNIT OF MEASURE, QTY, ESTIMATED UNIT APPROVED BUDGET OF THE CONTRACT, ESTIMATED TOTAL APPROVED BUDGET OF THE CONTRACT, OFFERED SPECIFICATION, QUOTED UNIT PRICE, TOTAL QUOTED PRICE, EVALUATION (Leave this space blank. For BAC/ Evaluators only)

1	Schlenk line and vacuum manifold setup	<p>Branded and brand new Specifications of vacuum pump: Displacement of 5.6 CMH; Stages:2; Type of oil: H/C or PFPE; Inlet connection: NW25; Outlet connection: BSP; Motor power: 0.25kW, 220-240V single phase Includes: at least two volume's worth of oil</p> <p>Specifications of manifold set-up parts: One (1) single manifold with 4- port; 4mm double oblique hollow high vacuum greased stopcocks; Ports are fitted with two 14/20 inner joints and two 24/40 inner joints and are spaced 5" center to center. #15 o-ring joints for all 4 ports two AF-0210-01 #15 stoppers three AF-0070-05 #15 o-ring joint to serrated hose adapters. 2-pc design rugged trap assembly Might include a pump.</p> <p>At least one (1) year warranty on parts and services</p>	lot	1	P166,000.00	166,000.00				
---	--	--	-----	---	-------------	------------	--	--	--	--

Cryogenic tank	<p>Branded and brand new Certifications/Compliance: CSA CE</p> <p>Able to maintain -180 °C sample environment with just 5cm of liquid nitrogen</p> <p>Dimensions: at least 55 cm (diameter) x at least 75 cm (height)</p> <p>Capacity: at least 65 L</p> <p>Type: Liquid nitrogen dewar</p> <p>Electrical requirements: 220V 50Hz</p> <p>Internal diameter: about 15 cm</p> <p>Rack Capacity: at least 6</p> <p>Static Holding time: at least 80 days</p> <p>Vial Capacity: at least 2000</p> <p>Includes at least 1 stainless steel rack designed for 5cm cryogenic boxes</p> <p>MIGHT include: at least 1 liquid nitrogen measuring stick at least 1 pair of water proof (mid-arm, medium) cryogloves, at least 1 apron</p> <p>at least one (1) year warranty against manufacturing defects and preventive maintenance for at least twice a year at least 1 session of Training of personnel for handling and care of cryogenic tank including certificate</p>	unit	1	P393,183.00	393,183.00					
TOTAL ABC					559,183.00	TOTAL QUOTED AMOUNT IN WORDS:				

Reviewed and Checked By:


 Mr. Eugenio D. Valencia
 BAC TWG

Please quote at your government price (including VAT) and state that the time within which you can make delivery. It will be appreciated if we can have your quotation in the office as

Ma. Jonah Girl H. Amoranto

BUYER/ END-USER

TERMS AND CONDITIONS:

1. Price quotation/s shall be valid for a period of at least (30) calendar days from the date of submission.
2. In order to assure that manufacturing defects shall be corrected by supplier, a warranty security shall be required from the contract awardee for a minimum period of **three (3) months**, in the case of EXPENDABLE SUPPLIES, or a
3. Delivery period within _____ calendar days.
4. Award of contract shall be made to the lowest quotation which complies with the technical specifications, and other terms and conditions stated herein
5. UPLB reserves the right to reject any or all offers as may be considered most advantageous to the University.
6. Any interlineations, erasure, or overwriting shall be valid only if they are signed or initiated by you or any of your duly authorized representative/s.

Requirements for Suppliers (GPPB Resolution No. 21-2017)

REQUIREMENTS	Shopping (Section 52)	Negotiated Procurement- Small Value Procurement (Section 53.9)	Negotiated Procurement- Lease of Real Property or Venue (Section 53.10)	Negotiated Procurement- Scientific, Scholarly or Artistic Work, Exclusive Technology and Media Services (Section 53.6)	Negotiated Procurement- Emergency (Section 53.2)
1 Mayor's/Business Permit	/	/	/	/	/
2 PhilGEPS Registration Number	/	/	/	/	/
3 Professional license/ CV (consulting services)		/		/	
4 PCAB License (for Infrastructure)		/			
5 Income / Business tax returns (except for government agencies as lessor)		FOR ABC'S ABOVE 500K	/	/	FOR ABC'S ABOVE 50K
6 Omnibus Sworn Statement		FOR ABC'S ABOVE 50K			FOR ABC'S ABOVE 500K
7 NFCC for Infrastructure with ABC above Ph500k		/			FOR ABC'S ABOVE 50K

After having carefully read and accepted your General Conditions, I/We quote you on the Item at prices noted above.

Name of the Company: _____

Address: _____ Tel. No. : _____

Signature over Printed Name of _____ Fax No. : _____

Position: _____ Email Address: _____

_____ Date: _____