



UNIVERSITY OF THE PHILIPPINES
LOS BANOS
 Los Banos, IV-A
 VAT Reg. TIN: 000-864-006-00004

UPLB BAC SECRETARIAT
 BY: JA DATE: 12/2/22
 DEC 06 2022 - ^{TIL} ₁₁₄₄
UPLB-RQ- 11-290-22-GOODS RECAPIVASC
DEADLINE OF SUBMISSION

Request for Quotation/ Bid Form (Technical Specifications)
BIOTECH-OVCAA-UPLB

UPLB-RQ-
 DEADLINE OF SUBMISSION: |

Suppliers Name: _____

Date: _____
 Fund Code: 9151030
 MOP: _____
 Contact No: 5362721
 Contact Person: Dr. Mannix S. Pedro

Please quote your lowest price on the item/s listed below, subject to the General Conditions below.

Note:

- Bidders shall provide correct and accurate information required in this form. All entries must be typewritten or in print and properly accomplished. Do
- Price quotation/s to be denominated in Philippine Peso shall include all taxes, duties, and/ or levies payable.
- Bidders must indicate the BRAND and MODEL NUMBER for equipment and its accessories or peripherals. Evidence shall be in the form of manufacturer's un-amended sale literature, unconditional statement of specification and compliance issued by the manufacturer and sample.
- Quotation through fax/email is acceptable. Winning bidder shall submit original signed RQ before issuance of Purchase order (P.O.).
- Quotations exceeding the Appoved Budget for Contract shall be rejected.
- Documentary requirements per Memorandum No. 03 Series of 2017 shall be attached upon submission of the quotation
- Others: _____

ITEM No.	GENERAL NAME OF THE ITEM	REQUIRED SPECIFICATIONS	UNIT OF MEASURE	QTY	ESTIMATED UNIT APPROVED BUDGET OF THE CONTRACT	ESTIMATED TOTAL APPROVED BUDGET OF THE CONTRACT	OFFERED SPECIFICATION <small>Suppliers must state here the detailed technical specifications of their offer against each of the individual parameters of each requirements</small>	QUOTED UNIT PRICE	TOTAL QUOTED PRICE	EVALUATION <small>(Leave this space blank. For BAC/Evaluators only)</small>
1	pH Meter	Supply, Delivery, Installaion and Commissioning of Brand New pH Meter A. Minimum specifications: 1.Type: Bench top pH meter 2. Measuring range: 0 to 14 pH 3. Calibration points: up to 5 4. pH Accuracy: $\leq \pm 0.003\text{pH}$ 5. Temperature Measurement range: -30.0 to 100.0°C 6. Temperature Accuracy: $\leq \pm 0.5^\circ\text{C}$ 7. Temperature Measurement Resolution: 0.1°C 8. ORP Range: $\pm 2000\text{ mV}$ 9. Includes glass body electrode, ATC probe, electrode holder and adaptor 10. Automatic / Manual temperature compensation 11. Built-in memory can store 500 data or more 12. Includes calibration buffers (atleast 3 points) B. Supplier qualifications:	2	un	70,000.00	140,000.00				

1. Supplier must be an authorized dealer of the brand that they offer.

2. Supplier must have trained personnel available locally for technical support.

C. Warranty and delivery period:


1. One (1) year warranty on parts and 2 years service.

2. Delivery period of ≤ 60 days after receipt of NTP.


Total abc 140,000.00

TOTAL QUOTED AMOUNT IN WORDS:

Reviewed and Checked By:


Mark Anthony Badua
BAC TWG

Please quote at your government price (including VAT) and state that the time within which you can make delivery. It will be appreciated if we can have your quotation in the office as soon as possible on or before the deadline stated herein.


Dr. Mannix S. Pedro
End User

TERMS AND CONDITIONS:

- Price quotation/s shall be valid for a period of at least (30) calendar days from the date of submission.
- In order to assure that manufacturing defects shall be corrected by supplier, a warranty security shall be required from the contract awardee for a minimum period of **three (3) months**, in the case of EXPENDABLE SUPPLIES, or a **minimum of one (1) year warranty and two (2) to three (3) years extended**
- Delivery period within 60 calendar days.
- Award of contract shall be made to the lowest quotation which complies with the technical specifications, and other terms and condition stated herein
- UPLB reserves the right to reject any or all offers as may be considered most advantageous to the University.
- Any interlineations, erasure, or overwriting shall be valid only if they are signed or initiated by you or any of your duly authorized representative/s.

Requirements for Suppliers (GPPB Resolution No. 21-2017)

REQUIREMENTS	Shopping (Section 52)	Negotiated Procurement- Small Value Procurement (Section 53.9)	Negotiated Procurement- Lease of Real Property or Venue (Section 53.10)	Negotiated Procurement- Scientific, Scholarly or Artistic Work, Exclusive Technology and Media Services (Section 53.6)	Negotiated Procurement- Emergency (Section 53.2)
1 Mayor's/Business Permit	/	/	/	/	/
2 PhilGEPS Registration Number	/	/	/	/	/
3 Professional license/ CV (consulting services)		/		/	
4 PCAB License (for Infrastructure)		/			
5 Income / Business tax returns (except for government agencies as lessor)		FOR ABC'S ABOVE 500K	/	/	FOR ABC'S ABOVE 50K
6 Omnibus Sworn Statement		FOR ABC'S ABOVE 50K			FOR ABC'S ABOVE 500K
7 NFCC for Infrastructure with ABC above Ph500k		/			FOR ABC'S ABOVE 50K

After having carefully read and accepted your General Conditions, I/We quote you on the Item at prices noted above.

Name of the Company: _____

Address: _____

Signature over
Printed Name of
Representative:

Position: _____

Tel. No. : _____

Fax No. : _____

Email Address: _____

Date: _____