

UNIVERSITY OF THE PHILIPPINES
LOS BANOS
Los Banos, IV-A
VAT Reg. TIN: 000-864-006-00004

UPLB BAC SECRETARIAT
BY: [Signature] DATE: 6-16-23
JUN 20 2023 5pm
UPLB-RQ- 6-191-23-RES
DEADLINE OF SUBMISSION

Request for Quotation/ Bid Form (Technical Specifications)
CENTER FOR AGRI-FISHERIES AND BIOSYSTEMS MECHANIZATION (BIOMECH)

UPLB-RQ-

DEADLINE OF
SUBMISSION:

Suppliers Name: _____

Date: May 15, 2023
Fund Code: NS-B31-32
MOP: SHOPPING
Contact No: 0960-467-9332
Contact Person: Maria Celeste P. Aquino
ciecielsaquino@gmail.com

Please quote your lowest price on the item/s listed below, subject to the General Conditions below.

Note:

- Bidders shall provide correct and accurate information required in this form. All entries must be typewritten or in print and properly accomplished. Do not
- Price quotation/s to be denominated in Philippine Peso shall include all taxes, duties, and/ or levies payable.
- Bidders must indicate the BRAND and MODEL NUMBER for equipment and its accessories or peripherals. Evidence shall be in the form of manufacturer's un-amended sale literature, unconditional statement of specification and compliance issued by the manufacturer and sample.
- Quotation through fax/email is acceptable. Winning bidder shall submit original signed RQ before issuance of Purchase order (P.O).
- Quotations exceeding the Approved Budget for Contract shall be rejected.
- Documentary requirements per Memorandum No. 03 Series of 2017 shall be attached upon submission of the quotation
- Others: _____

ITEM No.	GENERAL NAME OF THE ITEM	REQUIRED SPECIFICATIONS	UNIT OF MEASURE	QTY	ESTIMATED UNIT APPROVED BUDGET OF THE CONTRACT	ESTIMATED TOTAL APPROVED BUDGET OF THE CONTRACT	OFFERED SPECIFICATION Suppliers must state here the detailed technical specifications of their offer against each of the individual parameters of each requirements	QUOTED UNIT PRICE	TOTAL QUOTED PRICE	EVALUATION (Leave this space blank. For BAC/ Evaluator's only)
1	Color Meter	Branded and brand new color meter Illuminating/viewing geometry: 8°:d (8°illumination angle/diffuse viewing: specular component included (conforms to DIN 5033 Teil 7, JIS Z 8722 condition d, ISO 7724/1, CIE No. 15, ASTM E 116 Detector: Silicon photocells (6 pcs.), Display range: L*: 1 to 100 Light source: Pulsed xenon lamps, Measurement time: Approx. 1 sec Battery performance: Approx. 2,000 measurement at 10-sec. intervals when using alkaline-manganese batteries Measurement area: Approx. ø8 mm, approx. ø5 mm (optional accessory (*Optional ø5 mm mask does not conform to DIN 5033 Teil 7 and CIE N) Repeatability: Standard deviation within ΔE*ab 0.1 (when a white calibration plate is measured 30 times at 10-second intervals) Display languages: English, Simplified Chinese, and Japanese Interface: USB2.0, Observer: 10" Standard Observer, Illuminant: D65 Display data: Color difference, average (up to 10 times), Pass/fail judgment Color different formula: Δ (L*a*b*), Δ (L*C*H*), Δ E*ab (CIE 1976) Data memory: Target data, Samples data (up to 1,000 in total) Pass/fail item: Δ E*ab, Δ (L*a*b*), Δ (L*C*H*) Operation temperature/ humidity range: 0 to 40°C, relative humidity 85% or less (at 35°C) with no condensation Storage temperature/ humidity range: -20 to 40°C, relative humidity 85% or less (at 35°C) with no condensation Power: 4 AA-size alkaline dry batteries or	unit	1	170,000.00	170,000.00				

	nickel-metal-hydride rechargeable batteries, USB bus power or special AC Adapter Size (W x H x D): 66 x 158 x 85 mm, Weight: 420 g (without batteries)									
TOTAL ABC								170,000.00		
								TOTAL QUOTED AMOUNT IN WORDS:		

Reviewed and Checked By:

Donny D. Camus
BAC TWG 496-2023

Please quote at your government price (including VAT) and state that the time within which you can make delivery. It will be appreciated if we can have your quotation in the office as soon as possible on or before the deadline stated herein.

Maria Celeste P. Aquino
BUYER/END-USER

TERMS AND CONDITIONS:

- Price quotation/s shall be valid for a period of at least (30) calendar days from the date of submission.
- In order to assure that manufacturing defects shall be corrected by supplier, a warranty security shall be required from the contract awardee, for a minimum period of three (3) months, in the case of EXPENDABLE SUPPLIES, or a minimum of one (1) year warranty and two (2) to three (3) years extended warranty (if applicable) in the case of NON-EXPENDABLE SUPPLIES after acceptance by End-user of the delivered supplies.
- Delivery period within 30 calendar days.
- Award of contract shall be made to the lowest quotation which complies with the technical specifications, and other terms and condition stated herein.
- UPLB reserves the right to reject any or all offers as may be considered most advantageous to the University.
- Any interlineations, erasure, or overwriting shall be valid only if they are signed or initiated by you or any of your duly authorized representative/s.

Requirements for Suppliers (GPPB Resolution No. 21-2017)

REQUIREMENTS	Shopping (Section 52)	Negotiated Procurement- Small Value Procurement (Section 53.9)	Negotiated Procurement- Lease of Real Property or Venue (Section 53.10)	Negotiated Procurement- Scientific, Scholarly or Artistic Work, Exclusive Technology and Media Services (Section 53.8)	Negotiated Procurement- Emergency (Section 53.2)
1 Mayor's/Business Permit	/	/	/	/	/
2 PhilGEPS Registration Number	/	/	/	/	/
3 Professional license/ CV (consulting services)	/	/	/	/	/
4 PCAB License (for Infrastructure)	/	/	/	/	/
5 Income / Business tax returns (except for government agencies as lessor)		FOR ABC'S ABOVE 500K	/	/	FOR ABC'S ABOVE 50K
6 Omnibus Sworn Statement		FOR ABC'S ABOVE 50K			FOR ABC'S ABOVE 500K
7 NFCC for Infrastructure with ABC above Ph500k		/			FOR ABC'S ABOVE 50K

After having carefully read and accepted your General Conditions, I/We quote you on the item at prices noted above.

Name of the Company: _____

Address: _____

Signature over Printed _____

Name of Representative: _____

Position: _____

Tel. No.: _____

Fax No.: _____

Email Address: _____

Date: _____