



UNIVERSITY OF THE PHILIPPINES
LOS BANOS
Los Banos, IV-A
VAT Reg. TIN: 000-864-006-00004

Request for Quotation/ Bid Form (Technical Specifications)
INSTITUTE OF CROP SCIENCE

UPLB BAC SECRETARIAT
BY: FB DATE: 10-11-24
OCT 25 2024 10am
UPLB-RQ-10-468-24-RES
DEADLINE OF SUBMISSION

UPLB-RQ-
DEADLINE OF SUBMISSION:

Suppliers Name: _____

Date: 10/17/24
Fund Code: N8-407-21
MOP: Small Value Procurement
Contact No: 09163580427; lsbatolon@up.edu.ph
Contact: Lara Jaaziel S. Batalon/ ICropS

Please quote your lowest price on the item/s listed below, subject to the General Conditions below.

PL# 167663

Note:

- Bidders shall provide correct and accurate information required in this form. All entries must be typewritten or in print and properly accomplished. Do not leave
- Price quotation/s to be denominated in Philippine Peso shall include all taxes, duties, and/ or levies payable.
- Bidders must indicate the BRAND and MODEL NUMBER for equipment and its accessories or peripherals. Evidence shall be in the form of manufacturer's un-
- Quotation through fax/email is acceptable. Winning bidder shall submit original signed RQ before issuance of Purchase order (P.O.).
- Quotations exceeding the Approved Budget for Contract shall be rejected.
- Documentary requirements per Memorandum No. 03 Series of 2017 shall be attached upon submission of the quotation
- Others:

ITEM No.	GENERAL NAME OF THE ITEM	REQUIRED SPECIFICATIONS	QTY	UNIT OF MEASURE	ESTIMATED UNIT APPROVED BUDGET OF THE CONTRACT	ESTIMATED TOTAL APPROVED BUDGET OF THE CONTRACT	OFFERED SPECIFICATION Suppliers must state here the detailed technical specifications of their offer against each of the individual parameters of each requirements	QUOTED UNIT PRICE	TOTAL QUOTED PRICE	EVALUATION (Leave this space blank. For BAC/ Evaluator s only)
1	Polyacrylamide gel electrophoresis system	<p>Branded and Brand New Triple-Wide Vertical Electrophoresis System with Power Supply</p> <ul style="list-style-type: none"> - Tank Dimensions: approximately 19 (D) x 45 (W) x 12 (H) cm - Recommended Buffer Volume: Upper Reservoir - ~250 mL, Lower Reservoir - ~400 mL - Dual Unit (1-2 gels per run) - Glass Plate Dimensions: approximately 33cm (W) x 10cm (H) Electrophoresis cell with cooling chamber and coolant recirculation ports High-Throughput Mini-Gel screens up to 204 samples simultaneously Platinum electrodes Long-lasting silicone gaskets Power Input Requirement: 220V, 60Hz <p>1-300V Power Supply</p> <ul style="list-style-type: none"> - Net Weight: ~2.4 kg (5.3 lb) - Gross Weight: ~3.4 kg (7.5 lb) - Dimensions: approximately 320 (L) x240 (W) x73 (H) mm - Output Voltage: 1-300V, 1V steps, CC or CV - Output Current: 1-500mA, 1 mA steps - Timer: 1-1440 minutes (24 hours), 1 min. increment - Interface: 3-digit, red LED membrane keypad - Operating Conditions: Indoor use, 0-40°C. Relative Humidity: 0-90% noncondensing. Altitude: < 2000m - 24-month warranty on factory defaults <p>Safety cover with power leads White spring clamps 2 combs, 0.75mm thick x 31 teeth 2 sets plate spacers (0.75mm) 2 sets glass plates 2 lengths gasket 1 conversion plate for running one plate at a time</p>	1	set	200,000.00	200,000.00				
TOTAL						200,000.00				
								Total amount in words:		

Reviewed and Checked By:
Ariel U. Glorioso
BAC TWG

Please quote at your government price (including VAT) and state that the time within which you can make delivery. It will be appreciated if we can have your quotation in the office as soon as possible on or before the deadline stated herein.

Dr. Antonio G. Lalusin
BUYER/ END-USER

TERMS AND CONDITIONS:

- Price quotation/s shall be valid for a period of at least (30) calendar days from the date of submission.
- In order to assure that manufacturing defects shall be corrected by supplier, a warranty security shall be required from the contract awardee for a minimum period of **three (3) months**, in the case of EXPENDABLE SUPPLIES, or a **minimum of one (1) year warranty and two (2) to three (3) years extended warranty (if applicable)** in the case of NON-EXPENDABLE SUPPLIES after acceptance by End-user of the delivered supplies.

3. Delivery period within 30 calendar days.
4. Award of contract shall be made to the lowest quotation which complies with the technical specifications, and other terms and conditions stated herein.
5. UPLB reserves the right to reject any or all offers as may be considered most advantageous to the University.
6. Any interlineations, erasure, or overwriting shall be valid only if they are signed or initiated by you or any of your duly authorized representative/s.

Requirements for Suppliers (GPPB Resolution No. 21-2017)

REQUIREMENTS	Shopping (Section 52)	Negotiated Procurement- Small Value Procurement (Section 53.9)	Negotiated Procurement- Lease of Real Property or Venue (Section 53.10)	Negotiated Procurement- Scientific, Scholarly or Artistic Work, Exclusive Technology	Negotiated Procurement Emergency (Section 53.2)
1 Mayor's/Business Permit	/	/	/	/	/
2 PhilGEPS Registration Number	/	/	/	/	/
3 Professional license/ CV (consulting services)		/		/	
4 PCAB License (for Infrastructure)		/			
5 Income / Business tax returns (except for government agencies as lessor)		FOR ABC'S ABOVE 50K	/	/	FOR ABC'S ABOVE 80K
6 Omnibus Sworn Statement		FOR ABC'S ABOVE 50K			FOR ABC'S ABOVE 500K
7 NFCC for Infrastructure with ABC above Ph500k		/			FOR ABC'S ABOVE 90K

After having carefully read and accepted your General Conditions, I/We quote you on the item at prices noted above.

Name of the Company: _____

Address: _____

Signature over Printed

Name of Representative: _____

Position: _____

Tel. No. : _____

Fax No. : _____

Email Address: _____

Date: _____